

# bkSpecto Ultrasound System



## LEGAL MANUFACTURER

**BK Medical ApS**  
Mileparken 34  
2730 Herlev  
Denmark  
Tel.:+45 4452 8100  
www.bkultrasound.com  
Email: info@bkultrasound.com

The serial number on a BK Ultrasound product contains information about the year of manufacture.

### **BK Ultrasound Customer Satisfaction**

Input from our customers helps us improve our products and services. As part of our customer satisfaction program, we contact a sample of our customers a few months after they receive their orders. If you receive an email message from us asking for your feedback, we hope you will be willing to answer some questions about your experience buying and using our products. Your opinions are important to us. You are of course always welcome to contact us via your BK Ultrasound representative or by contacting us directly.

If you have comments about the user documentation, please write to us at the email address above. We would like to hear from you.

### **System Software**

- **NOT FAULT TOLERANT.** THE SOFTWARE IS NOT FAULT TOLERANT. THE MANUFACTURER HAS INDEPENDENTLY DETERMINED HOW TO USE THE SOFTWARE IN THE DEVICE, AND MS HAS RELIED UPON BK ULTRASOUND TO CONDUCT SUFFICIENT TESTING TO DETERMINE THAT THE SOFTWARE IS SUITABLE FOR USE.
- **EXPORT RESTRICTIONS.** You acknowledge that Windows 8 Embedded is of US-origin. You agree to comply with all applicable international and national laws that apply to Windows 8 Embedded, including the U.S. Export Administration Regulations, as well as end-user, end-use and country destination restrictions issued by U.S. and other governments. For additional information on exporting Windows 8 Embedded, see <http://www.microsoft.com/exporting/>
- The bkSpecto Ultrasound System is closed. Any modification of or installation of software to the system may compromise safety and function of the system. Any modification of or installation of software without written permission from BK Ultrasound will immediately void any warranty supplied by BK Ultrasound. Such changes will also void any service contract and result in charges to the customer for restoration of the original bkSpecto Ultrasound System.

### **Trademarks:**

DICOM is the registered trademark of the National Electrical Manufacturers Association for its standards publications relating to digital communications of medical information.

Microsoft and Windows are either registered trademarks or trademarks of Microsoft Corporation in the United States and/or other countries.

**bkSpecto= [Ref] 1300**

© 2018 BK Ultrasound

Information in this document may be subject to change without notice.

# Contents

---

<b>Chapter 1</b>	<b>General Information</b> .....	<b>5</b>
	Essential Performance .....	5
	Intended Use .....	6
	Modes of Operation .....	6
	Indications for Use .....	7
	Contraindications .....	7
<b>Chapter 2</b>	<b>Getting Started</b> .....	<b>9</b>
	The bkSpecto System .....	9
	Before You Start .....	10
	Height Adjustment .....	10
	Turning System On and Off .....	10
	Connecting Transducers .....	11
	Creating a User .....	11
	Starting an Exam .....	12
	Monitor and Touch Screen Display .....	14
	Basic Monitor Buttons .....	15
<b>Chapter 3</b>	<b>Safety Information</b> .....	<b>19</b>
	Safety Information .....	19
	Safety Symbols and Information on the Equipment .....	19
	General Safety Precautions .....	21
	Mechanical Safety .....	23
	Explosion Hazards .....	23
	Electrical Safety .....	24
	ESD Training .....	24
	Interference .....	24
	Electrical Noise .....	24
	Electromagnetic Interference .....	25
	RF (Radio Frequency) Interference .....	25
	Installation .....	26
	Power Cord and Fuses .....	26
	Connecting Other Equipment .....	27
	Network Connection .....	27
	Network Security .....	28
	Network Printing .....	28
	Connectors .....	29
	Video Output .....	30
	EMC Requirements .....	30
	Isolation of DICOM Network .....	31
	Wireless Networks .....	31
	Medical Equipment .....	34
	Non-Medical Equipment .....	34
	Battery Support System .....	34
	Computer Security .....	34

Printer .....	35
Service and Repair .....	35
Transducers .....	35
During an Examination .....	36
Checking the Date .....	36
Verifying the Transducer Type .....	36
Measurements .....	36
Puncture and Brachytherapy .....	37
3D .....	38
Acoustic Output .....	39
General .....	39
Monitor Display .....	40
Thermal and Mechanical Indices .....	40
Acoustic Output Measurement .....	41
Functions Affecting Acoustic Output .....	42
Default Acoustic Output .....	42
Clinical Measurements: Ranges and Accuracies .....	42
Geometric Measurements .....	44
Time Measurements .....	44
Doppler Measurements .....	45
<b>Chapter 4      Battery Support .....</b>	<b>47</b>
Before You Start .....	47
Imaging with Battery Support .....	47
Battery Location .....	47
Power Supply .....	47
Charging the Battery .....	47
Battery Status .....	47
Information Available on the Monitor .....	48
Battery Life .....	48
<b>Index .....</b>	<b>51</b>
<b>Appendix A      Warnings and Cautions Displayed on the System .....</b>	<b>53</b>

# Chapter 1

## General Information

---

This user guide is for all versions of the bkSpecto<sup>1</sup> ultrasound system.

**NOTE:** *Some of the functionality and options described in this guide may not be available with your version of the system.*

Before using the equipment, please make yourself familiar with the information in the accompanying user information documents. Some documents are printed. Make sure that you also read the transducer user guide and specifications for each transducer that you use.

Document	Information
System User Guide	Introductory information, safety information, getting started.
Quick Guide	User interface, basic operating instructions.
System Advanced User Guide	Information about advanced functions, glossary.
Product Data for system	Specifications for the system, including disinfection methods that can be used. Indications for use for each transducer that can be used with the system.
Technical Data (BZ2100)	Acoustic output data, clinical measurements (ranges and accuracies), factory default power levels and data about EMC (electromagnetic compatibility) for all transducers. Exam Type calculation formulas.
Care and Cleaning	Cleaning, disinfection, sterilization, checking, storing and disposing of BK equipment. Includes environmental limits.
Transducer User Guide	Specific instructions for the transducer and puncture attachments.
Product Data for each transducer	Specifications for the transducer, including disinfection methods that can be used.

*Table 1-1. User information documentation that accompanies the equipment.*

Improper use      Failure to follow safety instructions or use for purposes other than those described in the user manuals constitutes improper use.

### Essential Performance

The system can provide 2D and 3D ultrasound echo and flow imaging systems as an aid in diagnosis, data processing and -transfer, and guidance of puncture and biopsy.

The system can perform simple geometric measurements and calculations.

The system can guide biopsy- and puncture needles.

1. The bkSpecto has not been licensed by Health Canada.

The system is free from artefacts or distortion in the image or error of a displayed value, which can be attributed to a physiological effect and which may alter the diagnosis.

The system displays correct numerical values associated with the diagnosis to be performed.

The As Low As Reasonably Achievable (ALARA) principle is used and safety related indications (MI, TIS, TIB, etc) are displayed as worst-case values.

The system does not generate unintended or excessive ultrasound output or transducer surface temperature.

There is no unintended or uncontrolled motion of transducer assemblies intended for intra-corporeal use.

## **Intended Use**

The system is intended for diagnostic ultrasound imaging or fluid flow analysis of the human body, data processing and guidance of puncture and biopsy.

The system performs simple geometric measurements and calculations in the following areas:

- Urology

## **Modes of Operation**

- 2D (B-Mode) (including Tissue Harmonic imaging)
- M-Mode
- PWD Mode
- CFM Mode
- Power Doppler
- Elastography

## Indications for Use

The system is a diagnostic ultrasound imaging system used by qualified and trained healthcare professionals for ultrasound imaging, human body fluid flow analysis and puncture and biopsy guidance.

The clinical applications and exam types include:

- Abdominal
- Small Organ (also known as Small Parts)
- Transrectal

Indications for use are different for different transducers. The Product Data sheet for the system contains a table listing the indicated uses for each transducer that can be used with the system.

## Contraindications

- The bkSpecto ultrasound system is not intended for ophthalmic use or any use causing the acoustic beam to pass through the eye.






# Chapter 2 Getting Started

---



## The bkSpecto System



## Before You Start

Before you turn on the system, make sure that the installation has been approved by a qualified electrician or by hospital safety personnel. Plug the power cord into a grounded wall outlet and make sure that you can get to it easily when you need to turn off/unplug the system.

## Height Adjustment

The paddles underneath the keyboard lets you adjust the height and the angle of the system. Lift the right paddle to adjust the height , and the left paddle  to turn the system so you can operate it without the wheels getting in your way.

## Turning System On and Off

When you turn the system on or off, you must give the system enough time to save and recover open files and unsaved data. Otherwise, a serious system failure may occur that requires technical support.

The system has two power buttons, one on the monitor and one on the scanning engine:




Figure 2-1. The power button on the monitor and on the scanning engine, respectively.

This switch helps you to preserve the battery when the system is stored or otherwise not in use for a period.

### To turn the system on:

Press the power button *once*, then wait until startup screen disappears.

### To turn the system off:

Make sure system is running. Press the power button *once*.

Note that if you purchase the battery version of bkSpecto, it also has a **Battery Preserve Switch** located underneath the scanning engine:



Figure 2-2. Battery Preserve Switch. See location at “The bkSpecto System” .

## Connecting Transducers



Figure 2-3. Transducer sockets.

### To connect:

- 1 Insert transducer plug into socket with locking lever to the right.
- 2 Turn locking lever on socket to the left.

### To disconnect:

- 1 Freeze image.
- 2 Turn locking lever on socket to the right.
- 3 Remove plug from socket.



#### **WARNING** Exam-w2b

If, after beginning an exam without entering any Patient information, you want to save an image, you must verify that the system auto-created a properly configured Patient ID before exiting the exam. Whenever possible, BK recommends that you also enter a complete Patient Name.

## Creating a User

Only system administrators can create new users. See “Security Window” in the *bkSpecto Advanced User Guide*.


## Starting an Exam

The first 3 steps for imaging are:

- 1 Enter the patient information.
- 2 Select a transducer.
- 3 Select an exam type and a preset.

Do as follows:


- 1 Tap the **Patient** button on the touch screen.
- 2 Enter patient information. The **Patient ID** is filled in automatically with a date/timestamp, but you can change this to a relevant ID.



The screenshot shows the 'Patient' window interface. On the left is a vertical sidebar with buttons: PATIENT (highlighted in green), TRANSDUCER, PRESET, REVIEW, REPORT, END EXAM, and PAUSE EXAM. The main area is titled 'PATIENT DETAILS' and contains several input fields: PATIENT ID (20180118114236), LAST NAME, FIRST NAME, MI, DOB (MM/DD/YYYY), AGE, GENDER (dropdown), OPERATOR (DftUser), EXAM TYPE (dropdown, currently ABDOMEN), COMMENT, ACCESSION NUMBER, REFERRING PHYSICIAN, PREV EXAM DATE (MM/DD/YYYY), ADMITTING DIAGNOSIS, HEIGHT (CM), and WEIGHT (KG). At the bottom, there is a toggle for 'STORE IMAGES OF PATIENT DETAILS', and buttons for CLEAR, CANCEL EXAM, NEXT, and START EXAM.

Figure 2-4. Patient window.

- 3 If you select **Store Images of Patient Details**, an image of the patient details will be stored in the document browser and the review window.
- 4 Tap **Next** to select transducer. All connected transducers will be displayed in the **Transducer** window.



The screenshot shows the 'Transducer' window interface. The sidebar on the left has buttons: PATIENT, TRANSDUCER (highlighted in green), PRESET, REVIEW, REPORT, END EXAM, and PAUSE EXAM. The main area is titled 'Select the desired transducer' and displays two transducer options: '6C2' (highlighted with a green background) and '13L4W'. There are also two empty placeholder boxes. At the bottom, there are buttons for PREVIOUS, NEXT, and START EXAM.

Figure 2-5. Transducer window.

- 5 Tap **Next** to select **Exam Type** and **Preset** in the **Preset** window.




Figure 2-6. Preset window.

**6** Tap **Start Exam** to start the exam.

You can also start the exam immediately after entering **Patient Details**. Tap **Start Exam** and select transducer using the transducer control button. The system will use the default exam type and preset.




Figure 2-7. Start Exam

The exam ends when you tap **End Exam**:




Figure 2-8. End Exam.

# Monitor and Touch Screen Display



Figure 2-9. Monitor (clinical display).

- 1 Patient
- 2 Hospital name (logo)
- 3 Exam type/transducer
- 4 Image data
- 5 Measurement data
- 6 Document Browser
- 7 Wi-Fi and Storage Indicators



Figure 2-10. Touch screen.

## Basic Monitor Buttons



Figure 2-11. The basic monitor buttons on the touch screen



Button	Functionality
3D	Turn on 3D imaging. Double-tap to turn off.
PW	Turn on Pulsed Wave Doppler. Double-tap to turn off.
COL	Turn on Color mode. Double-tap to turn off.
2D	Turn on 2D imaging. Double-tap the 2D button to exit all modes and return to 2D mode only.
Home 	Tap the Home button to return to default settings for the current exam type.
TGC	Displays the TGC sliders.
Depth	Flick the bar to adjust image depth. Flick backward to increase depth and forward to decrease it. Tap either end to increase/decrease incrementally.
Focus	Flick the bar to adjust focus. Tap either end to adjust incrementally.
Gain	The Gain bar is located underneath the mode buttons. Slide this bar to adjust Gain.
Cine bar	Slide the Cine bar to scroll forwards and backwards. Tap either end to move one image at a time.
Optimize	Resets the TGC and optimizes the Doppler.
Dual View 	Toggles between the two images in dual mode. Only appears when dual mode is active.
Trackpad	Positions the pointer, measurement calipers and labels.
Measure	Generic measurements. Tap the trackpad once to place each caliper.
Calc	Specific measurements for the individual presets. Tap the trackpad once to place each caliper.
Label	Opens the Keyboard region with virtual keyboard, labels, bodymarks and arrows to label the clinical images.
Clear	Clears a measurement or a calculation. Only appears when Measure or Calc is active.
Back Trace	Erases backwards on a freehand drawing. Only appears when a freehand trace is active.
Print	Only appears when a printer is attached.
Freeze	Freezes/unfreezes live imaging.
Pointer	Displays a pointer on the monitor. Use the trackpad to move the pointer.
Store	Tap Store to save an image. When the image is stored, it will be displayed as a thumbnail at the bottom of the monitor.
Clip	Tap Clip to record a video clip. After recording, it will be displayed as a thumbnail at the bottom of the monitor.





Table 2-1. Basic Monitor Buttons explained.



- You select or deselect a button by tapping it.
- The buttons and window elements are highlighted in green when selected.

## Mode Button Backlight

The mode buttons are backlit according to their state:

	Mode selected. Note the line from the mode button to the Gain bar.
	Mode button enabled, but not selected. This mode is part of a combination mode.
	Mode not selected.
	Mode not available.

*Table 2-2. Mode button backlight*



## Chapter 3


# Safety Information


---

The system can be used for continuous operation, but imaging duration for individual patients must not exceed 60 minutes. We recommend, however, that you turn off the system at the end of each workday.

### Safety Information

This user guide contains cautions, warnings and other information about what you must do to ensure the safe and proper performance of the ultrasound system. You must also follow local government rules and guidelines at all times.

	<p><b>WARNING</b></p> <p>Warnings contain information that is important for avoiding personal injury.</p>
---	---

	<p><b>Caution</b></p> <p>Cautions contain information and instructions that must be followed to avoid damaging equipment, data, or software.</p>
---	--

**NOTE:** Notes contain information that you should be aware of.

### Safety Symbols and Information on the Equipment

Table 3-1 contains brief explanations of the symbols and information used to label the equipment. (Some labels in the table may appear on the transducer.)

The manufacturer disclaims all responsibility for the operating safety, reliability and performance of the equipment if these symbols and warnings are disregarded in any way.





Symbol	Name	Description
	Caution or Warning	Consult accompanying user guides when you encounter this sign on the instrument, to avoid reducing its safety.
	Follow instructions for use	Read the user guide or other instructions for important safety warnings.
	Consult instructions for use	Consult user guide or other instructions.
	Pushing prohibited	Do not use excessive force to push the system. Excessive force when pushing over uneven surfaces can cause the system to overbalance and tip.

Table 3-1. Symbols and information on the equipment.

















Symbol	Name	Description
	Keep hands clear	Show caution when you adjust the system monitor.
	SAFE WORKING LOAD	The weight in kilos of the system including transducers.
	Radio inside	The system has a built-in radio unit.
	Transmitter inside	The system contains a Dual Band Wireless Transmitter module with FCC-ID (US) and IC-ID (Canada).
	Manufacturer	Legal manufacturer.
	UL Classification for Canada and US	UL requirements are met for special conditions.
	Rx only	United States Federal law restricts this device to sale by or on the order of a physician.
	Potential Equalization	Terminal connected to the chassis. Should be connected to corresponding terminals on other equipment to eliminate potential differences.
	Type BF	BF: Isolated from ground. Maximum patient leakage current under <ul style="list-style-type: none"> <li>• Normal condition <math>\leq 100\mu\text{A}</math></li> <li>• Single-fault condition <math>\leq 500\mu\text{A}</math></li> </ul>
	Type BF	BF, defibrillator-proof.
	Type B	B: Maximum patient leakage current under <ul style="list-style-type: none"> <li>• Normal condition <math>\leq 100\mu\text{A}</math></li> <li>• Single-fault condition <math>\leq 500\mu\text{A}</math></li> </ul>
	Sealing	Dust- and immersion-protected according to EN 60529.
	Standby	Symbol on ON/Standby button on back of scanner unit and on the monitor – used to turn system on and off.
	ESD (electrostatic discharge)	Do not touch pins in connectors with this symbol unless you follow ESD precautionary procedures.
	Specified Radio Equipment	(On remote control UA2361) This equipment conforms to Japanese Radio Law regulations concerning frequency and power.
	WEEE waste	Within the EU, when you discard the equipment, you must send it to appropriate facilities for recovery and recycling.

Table 3-1. Symbols and information on the equipment. (continued)

Symbol	Name	Description
100-240V~50-60 Hz 300 VA F1, F2: T 4AH, 250 V	Mains	(Mains is placed at the bottom of the scanning engine). Indicates that system can be powered by 100-240 V ac, 50/60 Hz, Max power consumption 300 VA. Fuse cassette (F1, F2) contains two T 4AH, 250 V fuses.
<b>REF</b>	Reference number	1300. The marketing name for the system is <b>bkSpecto</b>
<b>SN</b>	Serial number	The serial number for the system.

Table 3-1. Symbols and information on the equipment. (continued)



## General Safety Precautions


The ultrasound system is designed and tested as Class 1 in accordance with EN/IEC 60601-1 (2012) (Part 1: General requirements for basic safety and essential performance) and EN 60601-2-37 (2007) (Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment).


The ultrasound system can optionally be internally powered by battery. All applicable EN/IEC 60601-1 tests have also been conducted on the ultrasound system while running on battery only.


The system also complies with ANSI/AAMI ES60601-1 (2005) and CAN/CSA C22.2 No.60601-1 (2014).


It fulfills the requirements for dust protection (IP20) for ordinary equipment specified in EN 60529.


 Physicians only	<p><b>Caution</b> Rx-c1</p> <p>United States Federal law restricts this device to sale by or on the order of a physician.</p>
 Proper Training	<p><b>WARNING</b> GS-w1</p> <p>To ensure safe and proper use of the equipment, before you attempt to use BK equipment, you should be trained in ultrasonography or be under the supervision of someone who is trained in ultrasonography. You should also be thoroughly familiar with the safe operation of your ultrasound system: read all the user documentation that accompanies it. No further training is required, but BK offers training in how to use the system. Consult your BK representative for information.</p>

 Equipment failure	<p><b>WARNING</b> GS-w2</p> <p>If at any time the system malfunctions, or the image is severely distorted or degraded, or you suspect in any way that the system is not functioning correctly:</p> <ul style="list-style-type: none"> <li>• Remove all transducers from contact with the patient.</li> <li>• Turn off the system. Unplug the system from the wall and make sure it cannot be used until it can be checked.</li> <li>• Do not try to repair the system yourself.</li> <li>• Contact your BK service representative or hospital technician.</li> </ul>
--	--

 Isolating the system	<p><b>WARNING</b> GS-w3</p> <p>The power supply cord connects the equipment to the line voltage. To isolate the equipment, you must unplug the power supply cord from the power source. Do this before you try to make any repairs to the system.</p>
---	---

 Spilled liquids	<p><b>Caution</b> S-c2</p> <p>The keyboard panel of the ultrasound system is <b>not</b> watertight. Be careful not to spill any liquids, gels or moist substances on the keyboard.</p>
--	--


 Condensation	<p><b>Caution</b> S-c3</p> <p>Large variations in temperature or humidity may cause water to condense inside the system. If this happens, the system may fail to operate properly. Always let the system come to room temperature before you plug it in.</p> <ul style="list-style-type: none"> <li>• Wait at least 2 hours after the system has been subjected to major changes in temperature or humidity.</li> <li>• If there is visible evidence of condensation, wait at least 8 hours.</li> </ul>
---	---


	<p><b>Caution</b> S-c4</p> <p>Never unplug the system from the wall while it is running. Turn off the system and wait for the light on the keyboard and the ON/Standby button to go out before unplugging.</p>
---	--


Before you use the equipment, make sure that all the safety requirements described in this chapter have been satisfied.


## Mechanical Safety

Mechanical failure or unintended use of ultrasound equipment can result in physical injury to patients or operators.


 Mechanical injury	<p><b>WARNING MS-w1</b></p> <p>Be careful to avoid the following potential sources of injury:</p> <ul style="list-style-type: none"> <li>• Parts of the body can be pinched by moveable parts of the equipment, such as the control panel.</li> <li>• Tilting the system can cause it to be unstable and injure someone.</li> <li>• Do not lean or sit on the control panel or any other part of the system. The control panel or monitor can break if subjected to heavy weights or impact.</li> </ul>
--	---

 All parts must be stable	<p><b>WARNING MS-w2</b></p> <p>When parts of the equipment can be mounted individually (for example, for use in an operating room) each part must be securely mounted to a stable support so that it does not tip, fall or come loose and injure someone.</p>
---	---


 	<p><b>WARNING MS-w3</b></p> <p>To avoid personal injury, be aware that the scanning engine can become hot after prolonged use.</p>
---	--


 Don't push too hard	<p><b>WARNING MS-w4</b></p> <p>To avoid injury and equipment damage, do not push the system too hard, especially when you roll the system over an uneven surface. Applying excessive force near the top could cause the system to overbalance and tilt.</p>
---	---

## Explosion Hazards

 Explosion hazards	<p><b>WARNING EH-w1</b></p> <p>The equipment is not designed to be used in potentially explosive environments. It should not be operated in the presence of flammable liquids or gases, or in oxygen-enriched atmospheres.</p> <p>There is a possible explosion hazard if the equipment is used in the presence of flammable anesthetic. The system should be placed at least 25 cm (10 inches) from the patient.</p> <p>The ultrasound system contains a lithium battery. Never remove or replace this battery. The lithium battery must not be removed except by a BK service representative.</p>
--	---

## Electrical Safety



 Do not use a power strip	<p><b>WARNING ES-w1</b></p> <p>Do not plug the equipment into an ordinary power strip. If the ground connection fails, this is dangerous because</p> <ul style="list-style-type: none"> <li>• the total leakage current for all the connected equipment can exceed the limits specified in EN/IEC 60601-1 (<i>Part 1: General requirements for safety</i>).</li> <li>• the impedance of the ground connection could exceed the limits specified in EN/IEC 60601-1.</li> </ul>
---	---

 Electrical shock	<p><b>WARNING</b> ES-w3</p> <p>You risk electrical shock if you try to get inside the equipment (other than opening a cover to access connectors described in the user guide). Do not allow anyone but qualified service personnel to service the equipment.</p>
---	--

## ESD Training

### The ESD Symbol


Anyone using the equipment must be able to recognize the ESD symbol and understand how to take the necessary precautionary procedures, as described in the caution below.

 ESD	<p><b>Caution</b> ESD-c1</p> <p>Do not touch pins in connectors that have the ESD symbol . Do not connect anything to them unless you follow these ESD (electrostatic discharge) precautionary procedures:</p> <ul style="list-style-type: none"> <li>• Discharge your body to ground before you touch the pins with your hand or a tool. For example, touch an unpainted metal part of the system cover.</li> <li>• You can use a wrist strap connected to the additional protective ground or potential equalization terminal on the system if that is more convenient.</li> </ul>
--	---

## Interference

The bkSpecto Ultrasound system is suitable for use in all establishments, other than domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.

### Electrical Noise


 Electrical noise	<p><b>WARNING</b> EN-w1</p> <p>Electrical noise from nearby devices such as electrosurgical devices – or from devices that can transmit electrical noise to the AC line – may cause disturbances in ultrasound images. This could increase the risk during diagnostic or interventional procedures.</p>
---	---

### Electromagnetic Interference

Medical electrical equipment requires special precautions regarding EMC (electromagnetic compatibility). You must follow the instructions in this chapter when you install the system and put it into service.

If the image is distorted, it may be necessary to position the system further from sources of electromagnetic interference or to install magnetic shielding.




 Other equipment nearby	<p><b>WARNING</b> EMC-w1</p> Do not use this equipment adjacent to other equipment. If you must place it next to or stacked with other equipment, verify that it operates normally there and neither causes nor is affected by electromagnetic interference.
---	--


EMC noise can reduce the usable image depth. Therefore, to avoid having to repeat an ultrasound examination, you must make sure beforehand that the ultrasound system can be used for the examination. Repeating an examination can be regarded as a potential risk that should be avoided, especially if the examination involves transducers used intracorporeally or transducers used for puncture.

## RF (Radio Frequency) Interference


Portable and mobile RF (radio frequency) communication equipment can affect the system, but the system will remain safe and meet essential performance requirements.

An ultrasound system intentionally sends /receives sound waves for the purpose of its operation. The transducers are very sensitive to frequencies within their signal frequency range (0.3MHz to 80MHz). Therefore RF equipment operating in this frequency range can affect the ultrasound image. However, if disturbances occur, they will appear as white lines in the ultrasound image and cannot be confused with physiological signals.

 Possible interference sources	<p><b>Caution</b> Inter-c1</p> Other equipment may interfere with the system, even if that other equipment complies with CISPR (International Special Committee on Radio Interference) emission requirements.
--	---

 Use specified equipment only	<p><b>Caution</b> Inter-c2</p> If you use accessories, transducers or cables with the system, other than those specified, increased emission or decreased immunity of the system may result.
---	--

## Installation

 <p>Installation safety requirement</p>	<p><b>WARNING</b> I-w1</p> <p>To ensure safe performance, a qualified electrician or hospital safety personnel must verify that the equipment is correctly installed and that it complies with the following safety requirements:</p> <ul style="list-style-type: none"><li>• Use only the original power supply cord. In the USA, this is fitted with a hospital grade three-prong grounded power plug. Never try to remove or change the plug on the power supply cord.</li><li>• All equipment must only be connected to a grounded AC power supply (or wall outlet) that meets EN/IEC/NEC requirements or applicable local regulations. The examination room's grounding system should be checked regularly by a qualified electrician or hospital safety personnel.</li><li>• Never use extension cords. The increased length of the cord will increase the resistance of the protective ground conductor and may increase the equipment's leakage current beyond an acceptable level.</li><li>• Keep power cords, sockets and plugs clean and dry at all times.</li><li>• Make sure that the power supply cord cannot be accidentally disconnected from the power source or the equipment.</li></ul>
--	--

Original power cords

If the original power cords are missing or damaged, you must order new ones from your local BK representative.

### Power Cord and Fuses

The system is delivered with a power cord that plugs into the power socket underneath the scanning engine.



Figure 3-1. Power socket and power cable. Note the label showing Mains.

The socket comes with a fuse cassette, located just above it. The cassette contains two fuses T 4AH, 250V.




Figure 3-2. Push tab to release fuse cassette.

## Potential Equalization






The potential equalization terminal  underneath the scanning engine is connected to the system chassis. It can be connected to corresponding terminals on other equipment to eliminate potential differences. Do NOT use it for additional protective grounding.



Figure 3-3. The terminals for potential equalization  are underneath the scanning engine.

## Connecting Other Equipment

For connection to other equipment, BK systems have a communication protocol on top of TCP/IP.

 Connection guidelines	<b>WARNING</b> C-w1 Follow the guidelines in EN/IEC 60601-1 when you connect the system to other equipment.
--	--

## Network Connection

BK's range of ultrasound systems comply with the DICOM standard for handling, storing, printing and transmitting information in medical imaging. DICOM includes a file format definition and a network communication protocol which facilitates the exchange of data between electronic medical systems.

For detailed information about:

- network requirements
- network configuration
- workflow between devices
- technical specifications
- safety specifications

see the DICOM conformance statement at

[www.bkultrasound.com/support/bk/resources/DICOM](http://www.bkultrasound.com/support/bk/resources/DICOM)

## Network Security

It is the responsibility of the on-site personnel or technician to maintain the IT-network and identify, analyze, evaluate and control new risks caused by a change in the network configuration.

If the applicable network connection does not meet the required characteristics of the IT-network, the following hazardous situations may occur:

- Corrupt patient data due to network errors, see Warning Exam-w3 on page 36
- System is unable to use the network due to faulty or overloaded network, see Warning GS-w1 on page 21
- System overloads the network causing other equipment to fail.

Network  
guidelines

**NOTE:** *If your system interacts with other equipment directly or indirectly you must ensure that your network is properly dimensioned and that critical equipment is placed on a separate network. Otherwise you could risk overloading the network and your equipment failing.*

## Network Printing

For printing on network printers, BK supports protocols PCL 5, PCL 6 and PS (Post Script).

## Connectors


The bkSpecto ultrasound system has three transducer sockets on the side of the system.

PC connectors for connecting the system to equipment such as approved printers and video equipment are located on the opposite side of the system. Do not use connectors that are not labeled.

Information about the correct cables to use is in Table 3-3.



*Transducer Sockets*



*System Connectors*

*Figure 3-4. Transducer sockets and system connectors.*

## System Connectors







Symbol	Connector	Additional Information
	Audio In	
	Audio Out	
	4 USB 3.0 connectors (A-type)	900mA current limit on each
	2 USB 2.0 connectors (A-type)	500mA current limit on each
	HDMI	To external monitor
	10/100/1000 Ethernet	LAN: 10/100/1000 LAN connector, RJ45

Table 3-2. System connectors.

## Video Output


Output signal is HDMI format only.

## EMC Requirements

To fulfill EMC requirements, cables attached to the system must be shielded and no longer than 5 m. The cables must be validated before use. Note that you cannot purchase these cables from BK Ultrasound.

Connector Name	Cable Type	Type and length
HDMI	HDMI	Shielded, 5m
Audio Out	Stereo, 3.5 mm jack	Shielded, 5m
Audio In	Stereo, 3.5 mm jack	Shielded, 5m
USB 1 Keyboard	USB, 3.0	Shielded, 5m
USB 2 Keyboard	USB, 3.0	Shielded, 5m
USB 3	USB, 2.0	Shielded, 5m
USB 4	USB, 2.0	Shielded, 5m
USB 5	USB, 3.0	Shielded, 5m
USB 6	USB, 3.0	Shielded, 5m
10/100/1000 Ethernet	Network, CAT6E	Shielded, 5m

Table 3-3. List of cables that fulfills EMC compliance.

 Use specified equipment only	<p><b>Caution Inter-c2</b></p> <p>If you use accessories, transducers or cables with the system, other than those specified, increased emission or decreased immunity of the system may result.</p>
---	---

## Isolation of DICOM Network

The system must not be galvanically connected to a computer network (DICOM®) that has not been isolated. If the network is not isolated, the system must be connected via a network isolator DP0925.

## Wireless Networks

The system can be connected to a wireless network for printing and archiving data.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- 1 This device may not cause harmful interference
- 2 This device must accept any interference received, including interference that may cause undesired operation

## Connecting to a Wireless Network

You must establish a secure wireless network at your hospital, clinic or institution, including a password for the network, before you can use the system's Wi-Fi for printing and archiving.



Figure 3-5. Position of the Wi-Fi symbol on the screen.

To connect to a wireless network, go to Settings, tap **Store/Network** and select the **Wi-Fi** tab. Tap **Configure Wifi**, select network on the monitor and tap **Connect** using the trackpad. Type in the network password. See also the section **Wi-Fi Tab** in the *bkSpecto Advanced User Guide*.




Figure 3-6. Wi-Fi connection window.

You only have to type in a password the first time you connect to a specific network. When you turn on the system an attempt is made to reconnect to the network that the system was connected to before it was turned off.



Weak vs.  
strong signal

The bars in the Wi-Fi logo represent the signal strength. The more lit bars in the logo, the stronger the signal.



No wireless  
connection

When the system is not connected to a wireless network, the Wi-Fi logo has a red 'X' over it.

An advanced Wi-Fi setting option is available on the system using Windows configurations. Always follow the security procedures that have been established for your hospital, clinic or institution, as well as national guidelines. Contact your BK service representative for more information.



<b>Additional characteristics</b>	
Dual band and Wi-Fi module	INTEL8260NGW
Operational frequency	2400-2483.5MHz 5150-5250MHz 5250-5350MHz 5470-5725MHz 5725-5875MHz
Channel spacing/bandwidth	13ch for 2.4GHz (5MHz overlap): 20/40MHz 24ch for 5 Ch (20MHz non overlap): 20/40/80MHz BT: 79 Ch (1MHz wide), BLE 3/37 Channels
RF output power	20dBm max.(2400-2483.5MHz) IEEE802.11 b/g/n mode 10 dBm max. (2400-2483.5MHz) Bluetooth/BLE 23 dBm max. (5150-5725MHz) IEEE 802.11 a/n/ac mode 14 dBm max. (5725-5875MHz) IEEE 802.11 a/n/ac mode
Type of modulation	Type: 2.4GHz: DSSS/OFDM/FHSS Type: 5GHz: OFDM BT 2.1 (+EDR), 3.0(+HS), 4.0 (BLE), 4.1, 4.2
FCC ID	PD98260NGU
Range	Wi-Fi depends on environment for the system. See below. Bluetooth by the transducer cable length (max. 3.5m). See Chapter 5.
Latency	Depends on network setup
Integrity	Full integrity of archiving operations
Security characteristics	Support for WEP, WPA, WPA2 and AES-CCMP. Enterprise encryption (802.1x) requires assistance from an authorized BK technician. Compliant with FIPS and FISMA.


Table 3-4. Additional characteristics.


When you transmit data over a wireless network, some special considerations apply. In particular, the network connections must be set up correctly. See Warning GS-w1 on page 21.

**NOTE:** *The system only supports one network at a time.*

If the system loses connection with the network while transmitting (for example, because it is moved out of range of the network during a transmission), the pending data is stored temporarily and re-transmitted when the connection is re-established.

For information on saving and printing using the DICOM protocol, see the DICOM chapter in the *Advanced User Guide* for the relevant system.


	<p><b>Caution:</b> Wifi-c1</p> <p>The network must be set up correctly so that data is sent to the correct location. Otherwise data can be lost or accessed by unauthorized people.</p>
---	---

	<p><b>Caution: Wifi-c2</b></p> <p>A safe encrypted protocol for data transmission, approved by the hospital, must be used. This is to prevent unauthorized people from getting access to the data.</p>
---	--

## Medical Equipment

If any other electrical equipment/accessory is connected to the system, the system *including* this equipment and/or accessory will become a medical system. Medical systems must comply with EN/EIC 60601-1, ANSI/AAMI ES60601-1 or CAN/CSA C22.2 No. 60601-1.

## Non-Medical Equipment

	<p><b>WARNING NME-w1</b></p> <p>Follow the guidelines in EN/IEC 60601-1.</p> <p>If you connect non-medical equipment (instruments that do not comply with safety requirements for medical equipment, such as a video monitor, video recorder, endoscopic camera control unit or other documentation device), this equipment must be placed outside the patient environment (1.5 m from the bed, for example). The equipment must fulfill the relevant EN standard or other applicable national or international standard.</p> <p>One of the following conditions must be fulfilled:</p> <ul style="list-style-type: none"> <li>• The system and other equipment are plugged into an external common isolation transformer to control the leakage current during a ground connection fault.</li> </ul> <p>or</p> <ul style="list-style-type: none"> <li>• The system is grounded with an additional safety ground connection (see "Potential Equalization" on page 27).</li> </ul> <p>If in doubt, contact your BK representative.</p>
---	---

## Battery Support System


If your system is battery powered, read the battery support chapter of this user guide.

Battery disposal


When you dispose of the batteries, you must follow national rules. Within the EU, you must send them to appropriate facilities for recovery and recycling.

## Computer Security

When BK Ultrasound systems are connected to a hospital network, the manufacturer does not take any responsibility for computer viruses from the network that may infect the system.


	<p><b>Caution: CS-c1</b></p> <p>You must perform a virus check on any external storage medium (USB device or DVD) to make sure that it is virus-free before you connect it to the system.</p>
---	---

## Printer


	<b>Caution:</b> Print-c1 The quality of a printed ultrasound image may vary, depending on the printer.
---	---


Note that images printed on the bkSpecto integrated printer are not to be used for diagnosis.


## Service and Repair

 Authorized personnel	<b>WARNING</b> SR-w1 Service and repair of BK electromedical equipment must be carried out only by the manufacturer or its authorized representatives. The manufacturer reserves the right to disclaim all responsibility, including but not limited to responsibility for the operating safety, reliability and performance of equipment serviced or repaired by other parties. After service or repairs have been carried out, a qualified electrician or hospital technician should verify the safety of all equipment.
---	---


## Transducers


 Electrical shock	<b>WARNING</b> T-w1 The transducer sockets contain terminals with 3.3V. Do not touch the patient while you are touching an uncovered socket.
--	---

 Type B transducers	<b>WARNING</b> T-w2 When using Type B (non-isolated) transducers, carefully check all electrical equipment within the patient area. Also, consider using additional protective grounding.
---	--

 Electrical burns	<b>WARNING</b> T-w3 Do not leave transducers in contact with the patient when using HF electrosurgical equipment.
---	--

BK Ultrasound transducers fulfill EMC requirements when they are outside as well as inside the patient's body.


 Surface temperature on array	<b>WARNING</b> T-w4 Do not turn the transducer on and allow it to scan into mid-air without ultrasound gel applied to the surface of the array. Doing so may cause the surface temperature on the array to heat up to 27 °C above room temperature (measured according to EN 60601-2-37 [3] ( <i>Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment</i> )). To avoid this, freeze the image when the transducer is not used for imaging.
---	---


 Creutzfeldt- Jakob disease	<b>WARNING</b> C-J-w1 Do not use a transducer for neurosurgical applications if the patient is suspected of having Creutzfeldt-Jakob disease. If a neurosurgical transducer has been used on a patient suspected of or diagnosed as being Creutzfeldt-Jakob positive, the transducer must be destroyed, following approved procedures for your hospital.
---	---


## During an Examination

### Checking the Date


Before you start imaging, verify that the date and time displayed on the monitor are correct.

 Date	<b>WARNING</b> Exam-w1 An incorrect date or time will make documentation of the image incorrect and may also cause some calculated values to be incorrect.
---	---

 Patient ID required	<b>WARNING</b> Exam-w2 You must enter a new patient ID or check that the system has entered a timestamp before you image a new patient. Otherwise, the documentation will not contain the correct patient identification, and you will not be able to capture images and clips. We recommend that you enter the complete name of the patient.
---	--


 Verify patient ID	<b>WARNING</b> Exam-w3 Verify that the patient name and ID are correct.
---	--


### Verifying the Transducer Type

 Type number displayed must match number on transducer	<b>WARNING</b> Exam-w4 Before you start to image, verify that the type number on the transducer matches the number displayed on the monitor. In case of any inconsistency, stop imaging, turn off the system, and contact your BK service representative.
---	--


### Measurements


Pay careful attention when you position cursors to make measurements on a scanned image or on a Doppler curve.


 Polygon measuring tool	<p><b>WARNING M-w1</b></p> <p>To prevent wrongful diagnosis, you must be aware that:</p> <ul style="list-style-type: none"> <li>When you use the polygon measuring tool, if the sides of the polygon intersect (as in forming a curve like a figure eight, for example), the area calculation is incorrect. In this case, the calculated area of the polygon is the area of the bigger loop minus the area of the smaller loop.</li> </ul>
---	--

 Using Doppler curves	<p><b>WARNING M-w2</b></p> <p>Drawings of Doppler curves, manual and automatic, are meant as tools for positioning cursors so that measurements based on the curves can be calculated automatically. The system has no facilities for checking whether the automatic measurements are reasonable. Curves drawn on very noisy spectra may lead to misplacement of measurement cursors. Make sure that measurement cursors are positioned so that the results are reasonable. If they are not, you must adjust the position of the cursors manually.</p>
---	--


## Puncture and Brachytherapy

 Verify transducer type number	<p><b>WARNING P-w1</b></p> <p>Before you start imaging, verify that the type number or name of the transducer and the type number or description of the puncture attachment you are using match the number displayed on the monitor. Also make sure that the needle guide is positioned correctly. If the numbers do not match, or if the needle guide position is not correct, the puncture line on the monitor may not correspond to the true puncture path in the tissue. In case of any inconsistency, stop imaging, turn off the system, and contact your BK service representative.</p>
--	---


 Verify puncture guide type number	<p><b>WARNING P-w2</b></p> <p>Verify that the type number of the puncture guide displayed on the monitor corresponds to the puncture guide that you are actually using. If the number is incorrect, the puncture line on the monitor may not correspond to the true puncture path in the tissue.</p>
--	--


 Watch the needle tip	<p><b>WARNING P-w4</b></p> <p>The puncture line on the image is an indication of the expected needle path. To avoid harming the patient, the needle tip echo should be monitored at all times so any deviation from the desired path can be corrected.</p>
---	--


**NOTE:** *If the image depth is set very low (to see tissue close to the transducer with high magnification), the needle tip echo can be outside the displayed image area. To see the needle tip in this case, zoom out so the full needle path is visible or pan the image to the side (to keep the high magnification).*


 Offset changes	<p><b>WARNING P-w3</b></p> <p>Changes you make to the offset of a programmable puncture guide or brachy matrix will affect ALL programmable puncture guides and brachy matrixes. This could lead to incorrect puncture lines or matrix positions for a different guide than the one you wanted to change.</p>
---	---

## Brachytherapy and Prostate Transperineal Biopsy

 Verify matrix type and coordinates	<b>WARNING B-w1</b> Verify that the matrix type and coordinates displayed on the monitor agree with the actual matrix template you are using.
---	--


 Verify user-defined matrix	<b>WARNING B-w2</b> If you create a user-defined matrix, it is your responsibility to verify that the matrix that appears on the monitor corresponds to the physical matrix you are using.
---	---


 Verify matrix alignment	<b>WARNING B-w3</b> Before you use the matrix for seed implantation or biopsy, check the matrix offset value to verify that it corresponds with the chosen matrix. Then check the matrix alignment.
--	--

 Verify matrix alignment	<b>WARNING B-w4</b> To avoid harming the patient, check the needle alignment (and recalibrate if necessary) before each use.
--	---

### 3D


Pay particular care to the following safety issues when operating the 3D system.

 3D measurements	<b>WARNING 3D-w1</b> Measurements obtained with the 3D system and used in diagnosis must be carefully and thoughtfully performed to ensure accurate quantitative assessment. Before you perform a calculation, make sure that all necessary calibrations and measurements are made.  If you suspect that the 3D system's calibration is inaccurate (that is, the measurements are not as expected), contact your BK service representative to check and confirm the system's proper operation.
--	---

 Untracked freehand	<b>WARNING 3D-w2</b> You cannot make accurate measurements on a 3D data set acquired using the untracked freehand method.
---	--

If you start to make a measurement on a 3D data set acquired using the untracked freehand method, the following warning appears on the monitor:

Warning on monitor


Symbol	Description
	Measurement will not be accurate.

## Acoustic Output


### General

Medical research has yet to prove whether or not ultrasound causes biological effects. Therefore, prudent use considerations require you to follow certain guidelines; see EN60601-2-37 (*Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment*).

### Prudent Use

 Exposure level	<b>WARNING AO-w1</b> To avoid tissue damage, always keep the exposure level (the acoustic output level and the exposure time) as low as possible.
---	--

- Image patients only when clinical reasons make it necessary.
- Keep exposure time as short as possible.
- Be careful to prepare the patient correctly so that you get the best possible image.
- Start imaging at a low acoustic output level (see “Thermal and Mechanical Indices” on page 40) and increase the level only as much as necessary to obtain a satisfactory image.
- If you switch from an application requiring high acoustic output levels (see “Functions Affecting Acoustic Output” on page 41), to one that requires lower levels (fetal imaging, for example), be sure to reset the levels before you image. (For example, start in 2D.)
- Take into account all the types of tissue that may be affected. For example, when imaging a breast, it may be appropriate to monitor the TI in bone rather than in soft tissue because the ribs will be subjected to ultrasound.

 Appropriate transducer	<b>WARNING AO-w2</b> To avoid tissue damage, always use the transducer best suited to the examination.
---	---

Acoustic output data for transducers used with the system are given in the Technical Data (BZ2100) that accompanies this user guide. The uncertainty level for each parameter is also listed. For definitions of the parameters, refer to the Food and Drug Administration (FDA) Guide as well as EN 60601-2-37 (*Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment*) and AIUM/NEMA standards.

The routes (or tracks) available for clearance by the FDA are well-defined. Track 3 is for diagnostic ultrasound systems that follow the Output Display Standard. Under Track 3, acoustic output will not be evaluated on an application-specific basis, but the maximum derated Spatial Peak–Temporal Average Intensity ( $I_{SPTA}$ ) must be  $\leq 720 \text{ mW/cm}^2$ , the maximum Mechanical Index (MI) must be  $\leq 1.9$ , and the maximum Thermal Index (TI) must be  $\leq 6$ . All BK Ultrasound transducers for use with the bkSpecto ultrasound system are Track 3.

## Monitor Display

The Mechanical Index (MI) and Thermal Index (TI) can be viewed in all imaging modes.

## Thermal and Mechanical Indices

The MI and TI indices are intended to allow users to implement the ALARA (As Low As Reasonably Achievable) principle using an indicator related to a potential bioeffect.

The full details of the indices are given in references EN60601-2-37 and AIUM/NEMA standard, but the formulas are given below.

### MI Formula

$$MI = \frac{P_{r0.3}(z_{sp})}{\sqrt{f_c}}$$

where the variables are defined in the table below.

Variable	Definition
$P_{r0.3}(z_{sp})$	Peak Rarefactional Pressure (MPa), derated by 0.3 dB/cm-MHz, measured at $z_{sp}$ , the point on the beam axis where pulse intensity integral ( $PII_{0.3}$ ) is maximum
$f_c$	measured center frequency (in MHz)

### TI Formula

$$TI = \frac{W_0}{W_{deg}}$$

where the variables are defined in the table below.

Variable	Definition
$W_0$	time-averaged acoustic power of the source or other power parameter (W)
$W_{deg}$	estimated power necessary to raise the temperature of the target tissue one degree Celsius ( $W/^\circ C$ )

Blood  
perfusion and  
TI

As a rule of thumb, the Thermal Index (TI) indicates the highest expected temperature increase in degrees Celsius. It is based on an average level of blood perfusion. The displayed TI may underestimate the temperature rise in poorly perfused tissues; you must take this into account when deciding on the maximum TI you will allow. Conversely, in areas with a rich perfusion of blood the temperature increase will be less than the displayed TI indicates.

Fever

A temperature increase of one degree Celsius increase in a patient with fever may cause complications in certain circumstances; it may be safer to delay the investigation.



## Acoustic Output Measurement

All values are measured in water according to the EN 60601-2-37 and AIUM/NEMA display standards. For some of the acoustic parameters, an estimated in situ derated value is given. This is derived assuming a tissue attenuation of 0.3 dB/(cm·MHz) when the estimated in situ derated value (I) is described by the following equation:

I formula  $I = I_w \exp(-0,069fz)$

where the variables are defined in the table below.

Variable	Definition
$I_w$	Intensity in water at the position where I is maximum
f	transducer frequency (in MHz)
z	distance (in cm) from the transducer face to the position where I is maximum

It should be stressed that the in situ values given are only applicable when there is attenuating tissue between the transducer face and the focal point.

## Functions Affecting Acoustic Output

The system has a control function that ensures that neither the  $I_{SPTA}$  nor MI nor TI value exceeds the maximum allowable value. When necessary, the system will reduce the output voltage and/or PRF (pulse repetition frequency) to the transducer to comply with requirements.

Some of the system functions can affect the acoustic output, as listed here. (Instructions for using these functions are given in the relevant sections of this user guide.)

- Sizing functions such as ROI (region of interest) – in general, smaller size results in higher acoustic intensity because the pulse repetition frequency (PRF) is higher or the ultrasound beam is more strongly focused.
- Focus – in general, strongly focusing the beam makes the acoustic intensity higher.
- Frame rate – higher frame rate results in higher acoustic intensity.
- Range – increasing the Doppler range increases the acoustic intensity by increasing the PRF.
- CFM Resolution – higher resolution increases the acoustic output.
- Color box size – narrowing the color box generally increases the acoustic output within it.

The user can set a Thermal Index limit. This will provide an upper limit for acoustic output.

## Default Acoustic Output

After the system has been turned off, the transducers will start in the default setup when the system is turned on again. The default setup may be factory-defined or defined by the user.

The factory-defined default setup values of acoustic output for each transducer are listed in the Technical Data (BZ2100).

These setups have been optimized to give the best compromise between low acoustic output and enough power to obtain the image features as quickly as possible. The factory default setup for all transducers is B-Mode to ensure the lowest acoustic output when you start imaging.

When you enter a new patient ID, by default the transducer setup will be reset to the factory setup.

## Clinical Measurements: Ranges and Accuracies

This section states the accuracies for measurements made using the BK range of ultrasound systems. A table containing accuracies for specific transducers can be found in the Technical Data (BZ2100) that accompanies this user guide.

The measurement accuracies are based on the assumption of “ideal” tissue, that is, a tissue characterized by a sound velocity of 1540 m/s. When making clinical measurements with ultrasound, errors may arise which are not taken into account in this section. For example,

- The sound velocity may vary from approximately 1450 m/s in fatty tissue to 1585 m/s in muscle. This can, in simple cases, give rise to errors of up to 6% for linear measurements. This inaccuracy may be further increased by refraction occurring at tissue boundaries.
- The user can introduce errors when using approximate formulas, when positioning the system’s calipers with respect to the ultrasound image and when outlining structures in the image.
- The ellipsoid volume approximation, described in this user guide, is only applicable when the cross section of the structure being studied approximates an ellipse (the circle being a special ellipse), and when the structure is roughly symmetrical about the selected axis of rotation

**NOTE:** *The choice of the axis of rotation is important for the calculation of the volume. A vertical axis gives a different volume than a horizontal axis.*

The minimum requirement is that the cross section outlined by the user should be convex. If the user draws a non-convex outline, an inaccuracy is introduced which is not taken into account in this section. In this case, the system calculates and displays the convex hull of the figure, that is, the smallest convex figure containing the non-convex figure outlined by the user (see Fig 3-7). The system bases the calculation of the ellipsoid volume on the convex hull.




Figure 3-7. The dotted line indicates the convex hull of the non-convex figure.

Measuring volume using a stepping method produces an approximation caused by the finite number of steps in the measurement. The user must always try to assess how large an inaccuracy is introduced by the selected step size, that is, the distance between organ cross sections.

## Geometric Measurements

### 2D Measurements

The geometric measurements performed by BK Ultrasound systems are distance, perimeter, area and ellipsoid volume. The accuracy of these measurements is influenced by the following factors:

- Transducer geometry
- Rounding of results
- Resolution of digital image memory

### 3D Measurements

The 3D volume is found by summing the marked area in the individual slices and multiplying by a factor that includes the distance between the slices and their relative orientation.

To obtain the accuracies listed in the table in the Technical Data (BZ2100), you must ensure that the calculation is based on contributions from at least 10 slices for very regular shapes and more for irregular shapes. It must also be possible to discriminate the boundary of the object from the surrounding tissue.

Volume accuracy

The volume accuracy in the table is given as a percentage of the captured 3D volume starting with the first slice and ending with the last slice that intersects the object of interest.

If the 2D ROI (region of interest) is set to be much larger than the object, the accuracy as a percentage of the object volume can be much worse (higher percent).

**NOTE:** *To ensure that the accuracy of your volume measurement is as high as possible, make sure that the object you are interested in fills the region of interest as much as possible.*

Distance and area accuracy

The accuracy of a distance measurement on a 3D image will never be better than 6%; the accuracy of an area measurement on a 3D image will never be better than 6%.

In the table in the Technical Data (BZ2100), the overall measurement accuracy for a full range measurement is given in the right-hand column for each measurement. The footnote below the table states the digital image resolution.

## Time Measurements

In M-mode and spectral Doppler mode, data is displayed along a time axis. It is possible to measure time differences. The accuracy for a time difference measurement is

- Rounded to the nearest: 0.01 s
- Accuracy:  $0.01 \times t$   
where  $t$  is the full time scale of the image field.

## Doppler Measurements

In measuring blood flow velocity it is assumed that the measured power spectral distribution of the Doppler signal equals the blood cell velocity distribution.

The measurement accuracy of blood flow velocity is heavily dependent on the angle  $\theta$  between the ultrasound beam and the velocity vector of the blood cells. To find the percentage accuracy, insert the angle:

$$\left( \frac{\cos \theta - \cos(\theta + 1,8)}{\cos \theta} \right) \times 100 + 0,5$$

### Examples:

0°	0.5%
15°	1.4%
30°	2.4%
45°	3.7%
60°	6.0%

The range of the blood flow velocity achievable at a 0° angle between the ultrasound beam and the velocity vector of the blood cells can be seen in the Technical Data (BZ2100).

If the blood velocity exceeds the selected velocity range, aliasing occurs, corresponding to an overload condition of the measurement system.

# Chapter 4

## Battery Support

---

This chapter is only relevant if your system has battery support.

The battery enables you to operate the system while it is not connected to an external electrical supply.

### Before You Start

Read the battery support warning in “Battery Support System” on page 34. Also read the rest of this chapter before you use the battery support system.

### Imaging with Battery Support

This is an overview of the steps for using the battery support system to power the imaging system.

- 1 Make sure the battery is charged.  
(If not, plug in the imaging system to use it or to charge the battery.)
- 2 Turn on the system.
- 3 When the battery is empty, you do not have to turn off the imaging system. Plug it into a power outlet to recharge the battery while you run on power from the normal power supply.

### Battery Location

The battery is located inside the system.

### Power Supply

#### Charging the Battery

The battery automatically begins charging when the system is plugged into the power outlet.

There is a battery charge display on the monitor. The battery will be fully charged after approximately 4 hours.

### Battery Status

While the system is operating, battery status is visible on the monitor (in the bottom right-hand corner).

## Information Available on the Monitor

The battery status indicator appears in the bottom right-hand corner of the monitor as a small 14 block e-ink display.



Figure 4-1. Battery level shown on the battery display. 100% battery capacity is left.

A message appears on the monitor when a battery reaches the end of its lifespan and needs replacing. For information about the disposal of depleted batteries, see page 34.

### Low Battery

When battery capacity is very low:

- A message appears on the monitor
- The on-screen battery indicator turns red





Figure 4-2. Low battery message.

The low-battery message informs you that the battery level is low and that the system will shut down when the system monitor clock matches the time stamp in the message (shown here as 15:11:23). If you want to keep using the system, plug it into a power outlet.

## Battery Life

With typical use (which includes imaging and freezing), a fully charged battery provides approximately 1 hour and 15 minutes of normal operation. With a fully charged battery pack, you can scan continuously for approximately 1 hour.

You can lengthen battery life by turning off the scanner using the battery preserve switch on the underside of the scanning engine when the system is not in use. This will prevent battery drain if the system is out of use for a longer period.



*Figure 4-3. The power switch underneath the scanning engine. See also “The bkSpecto System” on page 9.*

When a battery reaches the end of its lifespan, a replace battery message appears on the monitor at start-up.

On average, a battery will need to be replaced after approximately 1000 recharge and discharge cycles – the battery lifespan. Contact you BK Service technician for replacement of batteries.

For information about the disposal of depleted batteries, see page 34.





# Index

---

## Numerics

- 3D
  - measurements warning 38
  - warning about untracked freehand and measurements 39

## A

- accuracy
  - 2D geometric measurements 44
  - 3D distance 44
  - 3D measurements 44
  - 3D volume 44
  - clinical measurements explained 42
  - Doppler measurements 45
  - time difference measurements 44
- acoustic output
  - default 42
  - factors affecting 42
  - measurement 41
- ALARA principle 6, 40

## B

- battery
  - battery life 48
  - charging new battery 47
  - disposal 34
  - recycling 34
  - status 47, 48
- biopsy. See puncture.
- brachytherapy
  - warning about checking matrix 38
  - warning about user-defined matrixes 38
  - warning to verify matrix type and coordinates 38

## C

- caution or warning, symbol 19
- Cautions
  - about 19
  - displayed on the system 53
- center frequency 40
- charging battery 47
- connectors for other equipment
  - accessing 29
  - location 29
  - picture 29
- consult instructions for use (symbol) 19
- contraindications 7
- Creutzfeldt-Jakob warning 36

## D

- DICOM
  - conformance statement 27
  - connection if network is not isolated 31
  - network isolator 31
  - standard 27

- disposing of the system 21
- distorted image 25
- Doppler measurements
  - accuracies 45
  - warnings 36
- Doppler mode
  - fetal imaging 42

## E

- electrical noise 24
- electrical safety 26
- electrical shock 24
- electromagnetic compatibility. See EMC.
- electromagnetic interference 25
- EMC
  - attaching cables 30
  - data in BZ2100 5
  - noise, checking for 25
  - precautions 25
- equipment failure 22
- ESD
  - precautions 24
  - symbol 20, 24
  - training 24
- essential performance 5
- explosion hazards 23
- exposure level 39
- extension cord, do not use 26

## F

- fetal imaging, acoustic output settings for 42

## H

- HF electrosurgical equipment, warning 35

## I

- image distortion 22, 25
- indications for use 7
- installation 26
- intended use 6
- isolation from line voltage 22

## K

- keep hands clear, symbol 20

## L

- leakage current 24, 34
- liquids, do not spill on system 22

## M

- magnetic shielding 25
- manufacturer (symbol) 20
- measurements, Doppler warnings 36
- MI (Mechanical Index)
  - formula 40

modes of operation 6

## N

network

- guidelines 28
- isolator 31
- printing, protocols for 28

network security

- guidelines 28
- responsibility for 28

noise

- electrical 24
- EMC 25

non-medical equipment, connecting to system 34

## O

ON/Standby button

- symbol on 20

overbalance and tipping

- warning 23
- warning symbol 19

## P

patient ID

- verifying 36

Peak Rarefactional Pressure 40

potential equalization

- symbol 20
- terminal 27

power cord

- keep clean and dry 26
- original 26

power plug 26

power strip, do not use 24

power supply cord

- accidental disconnection 26
- use original 26

prostate transperineal biopsy

- warning about checking matrix 38
- warning about user-defined matrixes 38
- warning to verify matrix type and coordinates 38

prudent use 39

puncture

- warning about changing offset of programmable puncture guides or brachy matrixes 38
- warning to verify puncture guide number 37

pushing prohibited, symbol 19

## R

radio frequency interference. See RF interference.

radio inside, symbol 20

RF interference 25

rx only, symbol 20

## S

safety

- electrical 24, 26

transducer 35

sealing (symbol) 20

security, network. See network security.

service and repair 35

standby (symbol) 20

system

isolation from line voltage 22

malfunction 22

## T

Technical Data (BZ2100) 5, 40, 42, 44

Thermal Index limit

setting 42

Thermal Index limit, setting 42

TI (Thermal Index)

blood perfusion and 41

time measurements 44

tipping and overbalance

warning 23

warning symbol 19

training before using equipment 22

transducer sockets, warning about covering 35

transducer, electrical safety 35

transperineal biopsy

warning about checking matrix 38

warning about user-defined matrixes 38

warning to verify matrix type and coordinates 38

Type B (non-isolated) transducers, warning about 35

Type B (symbol) 20

Type BF (symbol) 20

Type BF, defibrillator-proof (symbol) 20

## U

UL (symbol) 20

UL Recognized Component for Canada and US, symbol 20

## V

virus

from network 34

## W

Warnings

about 19

displayed on the system 53

waste disposal symbol 21

WEEE waste (symbol) 21

weight, symbol 20

Wi-Fi

connecting to 31

correct setup and safe protocol required (cautions) 33

network characteristics 33

password 31

signal strength 32

using 31

wireless network. See Wi-Fi.

## Appendix A

### Warnings and Cautions Displayed on the System

---

This appendix contains a list of warnings and cautions that may appear on the system if you perform an undesirable action or e.g. type in a wrong setting value. Contact a BK service technician if you encounter a persistent problem that can not be solved by the suggested action(s) below.

This list covers all warnings and cautions on systems bkSpecto, bk3000, bk3500 and bk5000, so it may include warnings that are not relevant to your specific system.

The list also covers potential system and transducer malfunctions. Always contact a service technician if the system or the transducer malfunctions, see Warning GS-w2 on page 22.

<b>Warning text</b>	<b>Action (if necessary)</b>
Burn CD Failed	Try again.
Current date and time (%s) are before the scanner was last run (%s). Check your system clock.	
Error deleting file %1. Delete operation aborted.	
Error deleting folder %1. Delete operation aborted.	
Failed to clear the patient archive.	Please contact a service technician.
Failed to clear the patient archive because busy.	Cancel or wait until the DICOM Status Dialog Document List is empty before clearing the patient archive.
Failed to eject media.	
Failed to parse the prom received from the probe.	Try again.
Failed to read general settings from database.	Try again.
Failed to save bodymark catalog.	Try again.
Note that changing the time zone requires restarting the scanner.	Restart the system.
Parking of the transducer %s failed.	Try again.
Preparing CD/DVD Burning Failed.	
Probe prom read failed.	
Sensor data from probe is not received. Scanning is stopped.	

<b>Warning text</b>	<b>Action (if necessary)</b>
The packages could not be accessed.	
The probe could not be connected.	
The Probe FW is obsolete.	
The read probe prom has a wrong CRC.	
The read probe prom was empty.	
The system time is invalid and has been changed to %s %s.	The system time can be set in the setup menu.
The system time %s is invalid and setting it to %s failed.	The system time can be set in the setup menu.
There is an ongoing study in progress.	The ongoing study has to be stopped before the patient archive can be cleared.
There is not enough free space on the USB device. One or more documents were not copied, and the export is incomplete.	Please try again with a USB device with sufficient capacity.

<b>Caution text</b>	<b>Action (if necessary)</b>
%d text(s) has been truncated to %d characters.	
"%s" is not a unique name.	Type in a unique name.
"%s" is not a valid number.	Type in a valid number.
A 3D volume was not acquired because no mover is connected. Please connect a mover.	
A 3D volume was not acquired because the connected mover does not match the selected mover. The connected mover is %s and the selected mover is %s.	Make sure that connected mover matches the selected mover.
A curve cannot have itself as parent curve.	
A formula in the measurement contains an invalid device attribute "%s.%s".	Use valid device attribute.
A formula in the measurement contains an invalid result name "%s".	Use valid result name.
A formula in the measurement contains an undefined measurement tool "%s".	Use a defined measurement tool.
A measurement cannot depend on itself.	
Adjusting the guide's height.	
Adjusting the guide's width.	

<b>Caution text</b>	<b>Action (if necessary)</b>
All available pro packs for %s has been hidden	
An error occurred while the patient archiving database was being copied.	Try again.
Another scanner is using this network drive for archiving. This can cause problems. For independent backups, each scanner must use a separate subdirectory.	
Changes not in effect until transducer(s) has been reconnected.	
Corrupt user settings. System will restore your User Profile with default system settings.	
Click Yes to clear the patient archive. All patients in the patient archive will be lost. Click No to cancel.	
Could not create temporary subdirectory for outgoing mail data.	Try again.
Curve name, x-axis and y-axis are required. Do not leave the fields empty.	Fill out required fields.
Default LCD backlight has not been adjusted.	Adjust default LCD backlight.
Detected a transducer connection problem.	Please remove the transducer and connect it again.
Empty volume name is not allowed.	Type in volume name.
Engine configuration validation failed. One or more Engine item versions are not valid.	
Error opening database %s. %s	
Error. The file did not contain a user-defined puncture guide.	Include user-defined puncture guide.
Failed to acquire 3D volume. %s	
Failed to calculate expression "%s": %s.	Try again.
Failed to calculate measurement. Syntax error in formula.	Check formula.
Failed to clear patient archive. Please contact a service technician.	
Failed to clear the patient dialog cache. %s	
Failed to configure the server.	Try again.
Failed to connect to SMTP server "%s". (%s)	Try again.
Failed to copy file: "%s" To: "%s" %s	Try again.

<b>Caution text</b>	<b>Action (if necessary)</b>
Failed to copy.	Try again.
Failed to create Diagnostic Setup ID: %s.	Try again.
Failed to create new Pro Package %s.	Try again.
Failed to delete catalog in database.	Try again.
Failed to delete curve in database.	Try again.
Failed to delete measurement in database.	Try again.
Failed to delete the server.	Try again.
Failed to export table "%s". %s	Try again.
Failed to import license keys from ""%s"". It is not a valid license file.	
Failed to import table "%s". %s	Try again.
Failed to open file "%s". %s	Try again.
Failed to process 3D volume.	
Failed to read catalog from database.	Try again.
Failed to read deleted Presets.	Try again.
Failed to read key action definitions from database.	Try again.
Failed to read key assignments from database.	Try again.
Failed to read key definitions from database.	Try again.
Failed to read measurement list from database.	Try again.
Failed to restore factory default catalogs in database.	Try again.
Failed to restore factory default general settings.	Try again.
Failed to restore factory default measurement groups.	Try again.
Failed to save 3D volume.	
Failed to save all the DICOM Report Images.	
Failed to save configuration "%s" in database.	Try again.
Failed to save measurement in database.	Try again.
Failed to send SMTP message (%s).	Try again.

<b>Caution text</b>	<b>Action (if necessary)</b>
Failed to store catalog in database.	Try again.
Failed to store clip in file.	Try again.
Failed to store curve in database.	Try again.
Failed to store general settings in the database.	Try again.
Failed to store key assignments in database.	Try again.
Failed to store the curve "%s". Two curves cannot have the same Reference Name, and the imported curve's Reference Name "%s" is already used for the curve "%s".	Use a different reference name to store the curve.
Failed to store the DICOM Report.	
Failed to update Pro Package: %s.	Try again.
Failed to update the ICM file.	
Failed to validate database %s. Reverting to the latest backup from %s.	
Hard disk %s is %d%% full.	
Hard disk space on %s is critically low. Please clear some space on the hard disk or contact a service technician before proceeding.	
Hard disk space on %s is getting low.	Free up hard disk space.
Import after upgrade has already been performed once.	
Import of license keys from "%s" completed. The following license keys could not be imported: %s	Try again.
Invalid %s. Empty names are not allowed. Names must be shorter than %d characters. Names must not contain any of the following characters: %s	Fill out the required fields according to the on screen instructions.
License already in use.	
License key ""%s"" is invalid.	
Licenses missing: No available pro packs for %s has valid licenses	
Login error. Please try again.	
Name and description are required. Do not leave the fields empty.	Fill in name and description.
Name and formula are required. Do not leave the fields empty.	Fill in name and formula.

<b>Caution text</b>	<b>Action (if necessary)</b>
No available pro packages in database for transducer %s	
No catalogs available. Assign catalogs in setup window.	
No default propackage available for the %s transducer choose a default propackage	
No Preset name was specified. Nothing will be saved.	Specify a preset name.
One or more batteries are wearing out. Consider replacing the batteries.	
Only %d fields can be selected for the Patient window. Deselect some of the fields.	
Patient archive has been cleared. Restart the scanner to ensure data integrity.	
Pause/End of study is in progress.	
Please disconnect all transducers, restart the scanner, and start the import again.	
Please enter a number	
Please enter an SMTP Server Address. It must be entered as a name or an IP address.	
Please insert the archive disc labeled "%s" and try again.	
Pro Package %s could not be deleted because it would leave some transducers without a Pro Package.	
Staging area is not ready. Try again later.	
System is limited to %u captures.	
System restarts now.	
Temperature in probe is elevated. Adjust patient temperature and/or reduce probe heating policy.	
Temperature in probe is too high. Scanning is stopped and disabled until the temperature has come down.	
The '%s' curve could not be found in the database.	Try again.
The 3D volume is too big for the scanner memory. Please reduce the capture area or extent and try again.	



<b>Caution text</b>	<b>Action (if necessary)</b>
The AE title is invalid.	
The alias "%s" for the measurement "%s" is invalid. 1: Aliases cannot be empty. 2: Aliases must not contain any of the following characters: %s . 3: Aliases in each measurement must be unique.	Fill out the required fields according to the on screen instructions.
The archived document (%s) was not found on the network drive.	
The connected transducer cannot be used because there is a problem with its PROM. Please remove the transducer and contact a service technician.	
The curve definition is not compatible with this software version.	Upgrade software.
The curve definition must contain x and y values.	
The Display Controller Board has no ICM file.	
The document cannot be displayed because document data is corrupted	Contact BK service technician.
The document cannot be displayed.	Try again.
The document cannot be displayed. Currently there is no viewer available.	Install a viewer.
The entered license key has expired: %s.	
The entered license key is invalid.	Type in valid license key.
The field must not be empty.	
The file "\\\"%s\\" is too large to open. Max size is %d bytes.	If possible, reduce the file size.
The file does not contain a valid curve.	Use valid curve.
The file does not contain a valid measurement.	Use valid measurement.
The file size is too large to display on scanner	
The following licenses have expired: %s.	Install new license.
The formula exceeds the maximum length of %d characters.	Reduce length.
The guide's height is invalid.	Use valid height setting.

<b>Caution text</b>	<b>Action (if necessary)</b>
The guide's horizontal holes/cm is invalid.	Use valid horizontal holes/cm setting
The guide's horizontal offset is invalid.	Use valid horizontal offset setting.
The guide's name is invalid.	Use valid guide name.
The guide's vertical holes/cm is invalid.	Use valid vertical holes/cm setting.
The guide's vertical offset is invalid.	Use valid vertical offset setting.
The guide's width is invalid.	Use valid guide width setting.
The host name is invalid.	
The license could not be removed.	
The measurement cannot be edited. The template contains errors.	Contact service technician.
The measurement contains an invalid procedure name "%s".	Use valid procedure name.
The measurement could not be found in the database.	Try again.
The measurement definition is not compatible with this software version.	Upgrade software.
The measurement requires an unsupported measurement device type "%s".	Install supported measurement device type.
The measurement type is invalid.	Use valid measurement type.
The media was not burned: %s	
The name ""%s"" is already used.	
The name is invalid.	
The port number is invalid	
The procedure "%s" takes %d parameters.	
The report cannot be saved before a patient ID is entered.	
The report file cannot be generated: "%s"	Try again.
The same CD cannot contain both archived and non-archived documents.	
The scanner cannot communicate with the battery system.	If you want to run the scanner using battery power, we recommend that you contact a service technician.
The scanner is busy.	
The scanner is ready to enter service mode. The next time it starts, the PC will start in Windows mode.	Please press the standby switch to turn off the scanner.

<b>Caution text</b>	<b>Action (if necessary)</b>
The scanner is unable to communicate with other equipment through a COM port. %s	Use appropriate communication port(s). See <i>System User Guides</i> or contact service technician.
The scanner's hardware does not support %s.	
The selected curve is invalid.	Use valid curve.
The system has not finished sending the last mail.	Please try again later.
The system must be restarted for changes to take effect.	
The system must be restarted to fully implement the language change.	
The tool attribute %s is invalid.	Use valid tool attribute.
The transducer button %d is stuck on the transducer connected to connector %s.	
The two passwords you have typed are not identical.	
The USB Device cannot be safely removed.	Try again.
The user already exists.	Enter a different username.
The value "%s" (%s) contains the invalid character(s) "%s".	Type in valid characters.
The value "%s" (%s) is outside the legal range (%.2f - %.2f).	Keep value (s) within legal range.
The volume name is already used in the patient archiving system.	Please enter another name.
The volume name is not valid. It contains only illegal character(s).	Please enter the new volume name.
The volume name is too long. The name must be no longer than %d characters.	
The width and height must be numbers between %d and %d.	
There are no files to be burned to a CD.	
There is no mail receiver (SMTP To-address) configured.	Technical service must configure this service before it can be used.
There is no USB Device connected to the scanner.	
There was a problem saving the Diagnostic Setup %s.	Try again.

<b>Caution text</b>	<b>Action (if necessary)</b>
This scanner does not support CW Doppler. (This may be because of the power supply version.)	Try again or contact BK service technician.
Transducer "%s" cannot be used with this scanner.	Connect a compatible transducer. See <i>Product Data Sheet</i> for more details.
Transducer "%s" cannot be used with this scanner. A hardware upgrade may solve the problem.	Connect a compatible transducer. See <i>Product Data Sheet</i> for more details.
Transducer "%s" can't be used with this type of scanner.	Connect a compatible transducer. See <i>Product Data Sheet</i> for more details.
Transducer ""%s"" is disabled for this scanner.	Connect a compatible transducer. See <i>Product Data Sheet</i> for more details.
Unable to capture image.	Try again.
Unable to connect remote control.	
Unable to connect to the network drive.	Check server (UNC path), username and password.
Unable to connect to the network drive. %s	
Unable to create backup of database because of %s.	If the error persists, contact a service technician.
Unable to launch application ("%s")	Try again.
You must end the on-going exam before you can clear the patient archive.	End exam.
You must restart the scanner after you enable or disable the OEM interface.	Restart the system.
Your old password is not valid.	Create new password.



 BK Medical ApS, Mileparken 34, 2730 Herlev, Denmark. T +45 4452 8100 F +45 4452 8199



powered by analogic 

**BK Ultrasound**  
8 Centennial Drive  
Peabody  
MA 01960  
USA  
T +1 978-326-1300  
bkultrasound.com

**USA**  
**Sales & Service**  
BK Ultrasound  
8 Centennial Drive  
Peabody  
MA 01960  
USA  
T +1 978-326-1300  
F +1 978-326-1399  
bkultrasound.com

**Europe and Rest of World**  
**Sales, Service & Design Center**  
BK Ultrasound  
Mileparken 34  
2730 Herlev  
Denmark  
T +45 4452 8100  
F +45 4452 8199  
bkultrasound.com

**Asia**  
**Sales & Service**  
Analogic Medical Equipment  
(Shanghai) Co., Ltd.  
1377, Lan Dian Road  
Pu Dong New District  
Shanghai  
China 201132  
T +86 21 2089 0333  
bkultrasound.com