

Contents

Equipment 2

Preparations 2

1. Start up 4

2. Leakage test 6

3. Pressure levels 8

4. Trigger function 9

5. Upper pressure limit alarm 10

6. Tidal and minute volumes 12

7. Minute volume alarms 14

8. Check of "Neonate" range 18

9. Check tubings alarm 20

10. Apnea alarm 22

11. Safety valve 23

12. O₂ alarm 24

13. Gas supply system 26

14. Battery operation 30

15. Automode (SV 300 A) 33


16. Log sheet 34

A complete function check must be done after cleaning and after calibration (before the ventilator is connected to a patient).

If any malfunctions are detected during the function check, the ventilator must not be connected to patient before remedy of malfunction. The malfunction must be remedied by local technical staff trained by Siemens or a service technician from Siemens.


For those who do not have access to tubings and other accessories for adults a Neonatal kit, P/N 64 06 487 E380E, is available. The kit includes an alternative Operating Manual with Pre-use check, Calibration and Function check adapted for neonatal use.

Function check




Equipment

- 2 patient tubes.
- Y-piece.
- Test lung. Only a Siemens test lung shall be used.
- Gas supply: Air and O_2 .



Preparations


- Connect the ventilator to mains.



- Set the mode selector to "Stand by". Allow a warm-up period of at least 15 minutes.


Function check

300-F22E




- Connect gases.

- Open the lid on the patient unit.
- Set all knobs as shown.



SIEMENS

Servo Ventilator 300




SIEMENS

Servo Ventilator 300A


Function check

1. Start up


- Make sure the yellow light at "Ventilator off Battery charging" and the green light "Mains" are lit.




- Set the mode selector to "Stand by" and make sure:



- the **back-up alarm** (intermittent signals) is heard.
- all yellow lights are lit for a few seconds.
- all yellow and red lights in the "Alarms and messages" section stay lit for an additional moment.
- the caution signal is heard.



Function check




Alarms and messages

STAND BY

- the expiratory valve closes.
- the safety valve closes with a distinct click.
- the "Alarms and messages" display shows STAND BY.

These checks can be done separately by repeated switching between "Ventilator off Battery charging" and "Stand by".


Function check




2. Leakage test

Test for leakage and pressure transducer integrity


- Connect a patient tube between the inspiratory outlet and the expiratory inlet.



- Set the patient range selector to "Neonate".




- Set to "Pressure Control" mode.




- Make sure:
 - the diodes showing the actual pressure on the "Airway press." bargraph show the same value ± 5 cm H₂O. If not, see chapter Calibration. **Note:** The diodes can be hidden behind the diodes showing the set PEEP level.

Function check




- Keep "Pause hold" at "Exp." and make sure:




- the "End exp." display reading does not drop more than 10 cm H₂O during the exp. pause hold time (30 seconds).
- the diodes showing the actual pressure on the "Airway press." bargraph show the same value ± 5 cm H₂O. If not, see chapter calibration. **Note:** The diodes can be hidden behind the diodes showing the set PEEP level.

In case of leakage, check all connections in the ventilator.



- Release "Pause hold".




- Remove the patient tube.

Function check




3. Pressure levels

- Connect patient tubes, Y-piece and test lung.





- Set the patient range selector to "Adult".
- Set "Pressure Control Level above PEEP" to 30 cm H₂O.
- Set "PEEP" to 10 cm H₂O.




- Make sure the pressure indication on the "Airway press." bargraph equals the display "Peak" reading at the end of inspiration. The value must be in the range 38 – 42 cm H₂O.
- Make sure the pressure indication on the "Airway press." bargraph equals the display "End exp." reading at the end of expiration. The value must be in the range 9 – 11 cm H₂O.

4. Trigger function


- Set "Trig. sensitivity Level below PEEP" in the green range.



- Toward the end of expiration, quickly compress and release the test lung and make sure:
 - two yellow diodes at the lower right end of the "Airway press." bargraph flash once.
 - a breath is initiated.




- Set "Trig. sensitivity Level below PEEP" to -17 cm H₂O.




5. Upper pressure limit alarm

- Turn "Upper press. limit" **slowly** counter-clockwise and make sure upper pressure limit alarm is activated when "Upper press. limit" and the display "Peak" show the same value ± 2 cm H₂O.




- At the alarm activation, make sure:
 - the audible alarm is heard.
 - the inspiration stops and expiration starts.
 - the red light at "Airway pressure" in the alarm section lights up each time the alarm is activated.
 - the yellow light at "Airway pressure" is lit at the end of each alarm activation (when the red light at "Airway pressure" goes out).
 - the "Alarms and messages" display shows **Airway pressure too high**.
 - the upper pressure limit indication on the bargraph flashes.
 - the safety valve does not open. (If it opens, a distinct sound is heard and the PEEP-level drops to zero.)




- Set "Upper press. limit" to 60 cm H₂O.


Function check



- Set "Pressure Control Level above PEEP" to 0 cm H₂O.
- Set "PEEP" to 0 cm H₂O.



- Touch the touchpad at "Airway pressure" in the alarm section and make sure the display shows **Airway pressure too high**.




- Reset the alarm.


6. Tidal and minute volumes

Test of tidal and minute volumes and flow transducer integrity.


- Set to "Volume Control" mode.




- Adjust "CMV freq. b/min" so that the green display "Set freq. b/min" shows 20 b/min.



- Adjust "Volume" so that the green display "Tidal vol. ml" shows 375 ml.




Function check




30C0365E

- Make sure the green display "Minute vol. l/min" shows 7.5 ± 0.2 l/min.



30C0365E

- Wait a few breaths, then make sure:
 - the readings on the red display "Insp. tidal vol. ml" and the green display "Tidal vol. ml" correspond. Accuracy: ± 10 ml.
 - the readings on the red display "Exp. tidal vol. ml" and the green display "Tidal vol. ml" correspond. Accuracy: ± 10 ml.
 - the readings on the red display "Exp. minute vol. l/min" and the green display "Minute vol. l/min" correspond. Accuracy: ± 0.5 l/min.





30C0367E

- Also make sure the minute volume bargraph shows:
 - preset minute volume at the same value as the green display "Minute vol. l/min." Accuracy: ± 0.5 l/min
 - measured minute volume at the same value as the red display "Exp. minute vol. l/min". Accuracy: ± 0.5 l/min.


7. Minute volume alarms

Lower alarm limit

- Turn "Lower alarm limit" slowly clockwise and make sure:
 - the expired minute volume alarm is activated when the lower alarm limit indication passes the measured minute volume indication on the bargraph.




- At the alarm activation, make sure:
 - the audible alarm is heard.
 - the "Alarms and messages" display shows **Exp. minute volume too low**.
 - the red light at "Exp. minute volume" flashes.
 - the lower alarm indication on the bargraph flashes.
 - the lower alarm limit indication on the bargraph corresponds to the "Lower alarm limit" setting. Accuracy: ± 0.5 l/min.




- Turn "Lower alarm limit" to 0 l/min.

Function check




3000322E

- Make sure the yellow light at "Exp. minute volume" is lit.



3000321E

- Touch the "Exp. minute volume" touchpad and make sure the display "Alarms and messages" shows **Exp. minute volume too low**.




3C0-D68E

- Reset the alarm.


Function check

Upper alarm limit


- Turn "Upper alarm limit" slowly counter-clockwise and make sure:
 - the expired minute volume alarm is activated when the upper alarm limit indication passes the measured minute volume indication on the bargraph.




- At the alarm activation, make sure:
 - the audible alarm is heard.
 - the "Alarms and messages" display shows **Exp. minute volume too high**.
 - the red light at "Exp. minute volume" flashes.
 - the upper alarm limit indication on the bargraph flashes.
 - the upper alarm limit indication on the bargraph corresponds to the "Upper alarm limit" setting. Accuracy: ± 0.5 l/min.



- Turn "Upper alarm limit" to 60 l/min.




Function check




3000322E

- Make sure the yellow light at "Exp. minute volume" is lit.



3000323E

- Touch the "Exp. minute volume" touchpad and make sure the display "Alarms and messages" shows **Exp minute volume too high**.




300-D68E


- Reset the alarm.

8. Check of "Neonate" range


- Set the patient range selector to "Neonate."




- Make sure:
 - the yellow light "Neonate 1/10" is lit.



- the red light at "Technical" flashes.
- the display "Alarms and messages" shows **Overrange: Select pediatric**.




- the red "Insp. tidal vol. ml" display flashes.




- Check the "Upper alarm limit" and "Lower alarm limit" for expired minute volume as described in section 7.


Function check




- Turn "Volume" counter-clockwise until the display "Insp. tidal vol. ml" stops flashing. The green display "Tidal vol. ml" reading shall be below 40 ml.



- Set the patient range selector to "Adult."




- Set "Volume" so that the display "Tidal vol. ml" shows 375 ml.



- Reset all alarms.


9. "Check tubings" alarm

- Set to "Pressure Control" mode.
- Set "Pressure Control Level above PEEP" to 20 cm H₂O.
- Set "PEEP" to 5 cm H₂O.




3000324E

- Loosen the bacteria filter.




300-H60X

- Make sure:
 - the red light at "Technical" flashes
 - the display "Alarms and messages" shows **Check tubings**.
 - the safety valve opens for approx 5 seconds.




300-F60E


Function check



- Reconnect the bacteria filter.




- Set "Press. Control Level above PEEP" to 0 cm H₂O.
- Set "PEEP" to 0 cm H₂O.




- Reset the alarm.

10. Apnea alarm



- Set to "Volume Support" mode.




20 sec



- Wait for 20 seconds and make sure apnea alarm is activated.
- At the alarm activation, make sure:
 - the audible alarm is heard.
 - the "Alarms and messages" display shows **Apnea alarm**.
 - the ventilator changes from "Volume Support" to "Pressure Reg. Volume Control" mode (indicated by flashing yellow light at "Pressure Reg. Volume Control").




- Reset the alarm and make sure:
 - the ventilator switches back from "Pressure Reg. Volume Control" to "Volume Support."




- Set to "Volume Control" mode.

300-K57X



11. Safety valve

- Disconnect the patient tube from the inspiratory outlet.



300-F52E


- Cover the opening of the inspiratory outlet and make sure:
 - the audible alarm is heard.
 - the safety valve opens with a distinct click.
 - the display "Alarms and messages" shows **Airway pressure too high**.
 - the yellow light at "Airway pressure" is lit.

Do not activate the safety valve repeatedly since this may cause an overload of the electromagnet that controls the valve.



300-K58X

- Connect the patient tube to the inspiratory outlet.




300-D68E

- Reset the alarm.

12. O₂ alarm


Lower alarm limit

- Hold "Pause hold" at "Exp" and:




300-DS9E

- note the O₂ concentration value on the display "Alarms and messages." The reading shall be $40 \pm 3\%$ O₂.



3000330E


- Turn "O₂ conc. %" clockwise and make sure the "O₂ concentration" alarm is activated when the green display "O₂ conc. %" reading is $6 \pm 1\%$ **higher** than the noted value.



3000328E

- At the alarm activation, make sure:
 - the red light at "O₂ concentration" flashes.
 - the display "Alarms and messages" shows **O₂ conc too low**.

- Set "O₂ conc.%" to 40%.





3000329E


Function check

Upper alarm limit


- Hold "Pause hold" at "Exp." and:



- note the O_2 concentration value on the display "Alarms and messages." The reading shall be $40 \pm 3\% O_2$.



- Turn " O_2 conc. %" counter-clockwise and make sure the O_2 concentration alarm is activated when the green display " O_2 conc. %" reading is $6 \pm 1\%$ **lower** than the noted value.
- At the alarm activation, make sure:
 - the red light at " O_2 concentration" flashes.
 - the display "Alarms and messages" shows **O_2 conc too high**.




- Set " O_2 conc. %" to 40%.


13. Gas supply system




- Disconnect the O_2 supply and make sure Gas supply and O_2 concentration alarms are activated.



- At the alarm activation, make sure:
 - the audible alarm is heard.
 - the red light at "Gas supply" flashes.
 - the red light at " O_2 concentration" flashes.




- Touch the touchpad at "Gas supply" and make sure:
 - the "Alarms and messages" display shows **O_2 supply pressure too low.**
Air: X.X bar. O_2 : X.X bar.




- Mute the alarm and make sure:
 - the red display "Exp. minute vol. l/min" still shows the same value as the green display "Minute vol. l/min". Accuracy: ± 0.5 l/min.


Function check




- Connect O₂ supply.



- Wait a few breaths and then reset the alarms.




- Disconnect the air supply and make sure Gas supply and O₂ concentration alarms are activated.




- At the alarm activations, make sure:
 - the audible alarm is heard.
 - the red light at "O₂ concentration" flashes.
 - the red light at "Gas supply" flashes.

Function check




- Touch the touchpad at "Gas supply" and make sure:
 - the "Alarms and messages" display shows **Air supply pressure too low.**
Air: X.X bar. O2: X.X bar.




- Mute the alarm and make sure:
 - the red display "Exp. minute vol. l/min" still shows the same value as the green display "Minute vol. l/min". Accuracy: ± 0.5 l/min.


Function check



- Disconnect O₂ supply so that no gas is connected to the ventilator.
Make sure:
 - the safety valve opens with a distinct click.
 - the expiratory valve opens.




- Connect the air and O₂ supplies.




- Reset the alarms.

14. Battery operation

- Disconnect the ventilator from mains and make sure Battery alarm is activated.




3C0-D97X



3C0-D95E


- At the alarm activation, make sure:
 - the audible alarm is heard.
 - the red light at "Battery" flashes.
 - the display "Alarms and messages" flashes **BATTERY**.




3C0-D35E

- the yellow light at "Ventilator off Battery charging" and the green light "Mains" are not lit.

Function check




- Reset the alarm and make sure:
 - the caution signal starts.
 - the yellow light at "Battery" is lit.




- To check the internal battery, touch the touchpad at "Battery" and read the displayed text. (Normal value approx. 24 V.)

Function check


300037E



- Connect the ventilator to mains again.



- Make sure :
 - the yellow light at "Ventilator off Battery charging" and the green light "Mains" are lit.
 - the caution signal stops.





- the display "Alarms and messages" no longer flashes **BATTERY**.
- the yellow light at "Battery" is no longer lit.


300038E

15. Automode (SV 300A only)

- Set "Automode" to "On".
- Set "Trig. sensitivity Level below PEEP" in the green range.




- Compress and release the test lung twice to trigger a breath. Two diodes on the bargraph for airway pressure show that a breath is triggered.




- Make sure the ventilator switches to support mode. This is shown by the yellow light "Support".


Function check



- Do not trigger any more breaths and make sure the ventilator switches back to "Volume Control" after about 12 seconds. (The yellow light "Support" is no longer lit.)



- Set "Trig sensitivity Level below PEEP" to -17 cm H₂O.
- Set "Automode" to "Off"



- The Function check is now complete, set the mode selector to "Stand by", or, if the ventilator is to be connected to a patient, make the appropriate settings for that patient.

16. Log sheet

- Note on a log sheet that a function check has been performed.