

CCBRT Maintenance Workshop

Week 25

priority

done

out of workshop: Frank 16/6 – 18/6 Meeting with GTZ

Department	Equipment	Problem	To Be Done	Finished
Laboratory	Microscope	poor image	cleaning	
Eye Theatre	Autoclave	not heating	checking	
Eye Theater	UPS 1500	not working	repair, batteries changed, under test	
Ortho Theater 2	Autoclave	door leakage	gasket ordered (10/5/2009)	ordered
Eye Theatre	Gas supply system	leakage	→ work outsourced	
Theatre 3	Oxygen	regulator	change	Monday
Orth Workshop	Plaster saw	motor defective	repair	Thursday
Eye OPD	Aircon	cooling	Repair, fan rewinding	
Otho Theatre	AC	maintenance		Saturday
Eye	Cables	insulation(on top of pole)	change	Sat wk 27
Eye OPD	Tonometer	?	waiting for user to test	
Eye OPD	Microscope	eye pieces	ordered 11/6	ordered
Theatres	Mains cable/change over	power consumption	checking power consumption, goes to 150A	July
Workshop	Car	maintenance	to car workshop	July